

CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001716270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
ABD-Bandsaw01	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001716


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001716270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.11°C	-20.29°C	-20.29°C	0.17°C
27-Apr-17 12:39	9.84°C	9.79°C	9.79°C	0.05°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001664270417

Page 1 of 2

DATE OF ISSUE 27-Apr-17

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
ABD-Coldstore01	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001664


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001664270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.11°C	-20.00°C	-20.00°C	-0.11°C
27-Apr-17 11:32	10.06°C	9.81°C	9.81°C	0.24°C

The above calibration is applicable only to the relevant recorder and probe.

*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001722270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
ABD-Coldstore02	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001722


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001722270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.11°C	-20.12°C	-20.12°C	0.01°C
27-Apr-17 10:45	9.93°C	10.02°C	10.02°C	-0.10°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001717270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
ABD-Factory	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001717


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001717270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:17	-20.12°C	-20.29°C	-20.29°C	0.17°C
27-Apr-17 12:39	9.84°C	9.82°C	9.82°C	0.02°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001727270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
ABD-IntakeChill	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001727


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001727270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.11°C	-20.31°C	-20.31°C	0.20°C
27-Apr-17 11:35	10.04°C	9.74°C	9.74°C	0.31°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001723270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
ABD-MainChill02	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001723


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001723270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.12°C	-20.20°C	-20.20°C	0.08°C
27-Apr-17 11:20	9.88°C	9.93°C	9.93°C	-0.06°C

The above calibration is applicable only to the relevant recorder and probe.

*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001712270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
ABD-SalmonChill	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001712


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001712270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.11°C	-20.30°C	-20.30°C	0.19°C
27-Apr-17 11:36	10.03°C	9.53°C	9.53°C	0.50°C

The above calibration is applicable only to the relevant recorder and probe.
*Error comparing the reference equipment with the equipment under test.
**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001711270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
ABD-SalmonHeads	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001711


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001711270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.11°C	-20.30°C	-20.30°C	0.19°C
27-Apr-17 11:35	10.05°C	9.55°C	9.55°C	0.50°C

The above calibration is applicable only to the relevant recorder and probe.
*Error comparing the reference equipment with the equipment under test.
**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00002730270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
CSF-Coldstore01	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00002730


Environmental Conditions

Temperature	15°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00002730270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.11°C	-20.20°C	-20.20°C	0.08°C
27-Apr-17 12:28	9.92°C	9.63°C	9.63°C	0.29°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00002761270417

Page 1 of 2

DATE OF ISSUE 27-Apr-17

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
CSF-Coldstore02	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00002761


Environmental Conditions

Temperature	15°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00002761270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:17	-20.12°C	-20.23°C	-20.23°C	0.11°C
27-Apr-17 12:39	9.84°C	9.81°C	9.81°C	0.02°C

The above calibration is applicable only to the relevant recorder and probe.
*Error comparing the reference equipment with the equipment under test.
**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00002763270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
CSF-Coldstore03	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00002763


Environmental Conditions

Temperature	15°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00002763270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:17	-20.12°C	-20.26°C	-20.26°C	0.14°C
27-Apr-17 11:31	10.04°C	9.82°C	9.82°C	0.23°C

The above calibration is applicable only to the relevant recorder and probe.
*Error comparing the reference equipment with the equipment under test.
**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001726270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
DLS-Room1-1	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001726


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001726270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:34	-20.10°C	-20.10°C	-20.10°C	0.00°C
27-Apr-17 11:20	9.88°C	9.95°C	9.95°C	-0.07°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
 Studio 19
 50-54 St Pauls Square
 Birmingham, West Midlands
 B3 1QS

CERTIFICATE NUMBER 00001725270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
DLS-Room1-2	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001725


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001725270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:17	-20.12°C	-20.22°C	-20.22°C	0.10°C
27-Apr-17 11:32	10.06°C	9.79°C	9.79°C	0.26°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001724270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
DLS-Room2-1	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001724


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001724270417

Page 2 of 2

Calibration Results


	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.11°C	-20.16°C	-20.16°C	0.05°C
27-Apr-17 11:32	10.05°C	9.83°C	9.83°C	0.23°C

The above calibration is applicable only to the relevant recorder and probe.
*Error comparing the reference equipment with the equipment under test.
**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001720270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
DLS-Room2-2	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001720


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001720270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.11°C	-20.17°C	-20.17°C	0.06°C
27-Apr-17 11:32	10.06°C	9.80°C	9.80°C	0.27°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001719270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
DLS-Room3-1	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001719


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001719270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.11°C	-20.20°C	-20.20°C	0.09°C
27-Apr-17 11:32	10.06°C	9.65°C	9.65°C	0.41°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001718270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
DLS-Room3-2	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001718


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001718270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.11°C	-20.10°C	-20.10°C	-0.01°C
27-Apr-17 11:31	10.04°C	9.74°C	9.74°C	0.31°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001715270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
DLS-Room4-1	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001715


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001715270417

Page 2 of 2

Calibration Results


	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:26	-20.10°C	-20.11°C	-20.11°C	0.01°C
27-Apr-17 11:32	10.06°C	9.66°C	9.66°C	0.40°C

The above calibration is applicable only to the relevant recorder and probe.
*Error comparing the reference equipment with the equipment under test.
**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001710270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
DLS-Room4-2	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001710


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001710270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:17	-20.12°C	-20.17°C	-20.17°C	0.05°C
27-Apr-17 11:35	10.05°C	9.63°C	9.63°C	0.42°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001679270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
DLS-Room4-3	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00002254
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1296
Probe Certificate Issue Date	12-Jul-16
Probe Certificate Reference Number	00002254110716
Equipment Code(s)	CI0001, CI0005

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001679


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001679270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 14:43	-20.05°C	-19.94°C	-19.94°C	-0.11°C
27-Apr-17 11:00	9.93°C	10.05°C	10.05°C	-0.12°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
 Studio 19
 50-54 St Pauls Square
 Birmingham, West Midlands
 B3 1QS

CERTIFICATE NUMBER 00001677270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
DLS-Room4-4	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00002254
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1296
Probe Certificate Issue Date	12-Jul-16
Probe Certificate Reference Number	00002254110716
Equipment Code(s)	CI0001, CI0005

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001677


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001677270417

Page 2 of 2

Calibration Results


	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 14:41	-20.02°C	-19.96°C	-19.96°C	-0.06°C
27-Apr-17 11:00	9.93°C	10.12°C	10.12°C	-0.19°C

The above calibration is applicable only to the relevant recorder and probe.
*Error comparing the reference equipment with the equipment under test.
**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001665270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
DLS-Room4-5	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00002254
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1296
Probe Certificate Issue Date	12-Jul-16
Probe Certificate Reference Number	00002254110716
Equipment Code(s)	CI0001, CI0005

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001665


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001665270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 14:47	-20.11°C	-19.95°C	-19.95°C	-0.17°C
27-Apr-17 11:00	9.93°C	10.10°C	10.10°C	-0.17°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001663270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
DLS-Room4-6	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00002254
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1296
Probe Certificate Issue Date	12-Jul-16
Probe Certificate Reference Number	00002254110716
Equipment Code(s)	CI0001, CI0005

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001663


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001663270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 14:41	-20.02°C	-19.96°C	-19.96°C	-0.06°C
27-Apr-17 11:00	9.93°C	10.10°C	10.10°C	-0.17°C

The above calibration is applicable only to the relevant recorder and probe.
*Error comparing the reference equipment with the equipment under test.
**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001670270417

Page 1 of 2

DATE OF ISSUE 27-Apr-17

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
EQY-Blast01	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001670


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001670270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:17	-20.12°C	-20.32°C	-20.32°C	0.20°C
27-Apr-17 11:35	10.05°C	9.56°C	9.56°C	0.49°C

The above calibration is applicable only to the relevant recorder and probe.
*Error comparing the reference equipment with the equipment under test.
**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001678270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
EQY-Blast02	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001678


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001678270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.11°C	-20.28°C	-20.28°C	0.17°C
27-Apr-17 11:32	10.06°C	9.70°C	9.70°C	0.36°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001681270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
EQY-Blast03	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001681


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001681270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:17	-20.12°C	-20.32°C	-20.32°C	0.20°C
27-Apr-17 11:20	9.88°C	9.75°C	9.75°C	0.13°C

The above calibration is applicable only to the relevant recorder and probe.
*Error comparing the reference equipment with the equipment under test.
**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001683270417

Page 1 of 2

DATE OF ISSUE 27-Apr-17

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
EQY-Blast04	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001683


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001683270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.11°C	-20.17°C	-20.17°C	0.05°C
27-Apr-17 11:32	10.05°C	9.79°C	9.79°C	0.27°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00003051270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
EQY-Blast05	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00003051


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00003051270417

Page 2 of 2

Calibration Results


	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:17	-20.12°C	-20.12°C	-20.12°C	0.00°C
27-Apr-17 11:34	10.07°C	9.84°C	9.84°C	0.23°C

The above calibration is applicable only to the relevant recorder and probe.
*Error comparing the reference equipment with the equipment under test.
**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001690270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
EQY-Blast06	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001690


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001690270417

Page 2 of 2

Calibration Results


	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.11°C	-20.13°C	-20.13°C	0.02°C
27-Apr-17 11:32	10.06°C	9.64°C	9.64°C	0.41°C

The above calibration is applicable only to the relevant recorder and probe.
*Error comparing the reference equipment with the equipment under test.
**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001701270417

Page 1 of 2

DATE OF ISSUE 27-Apr-17

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
EQY-Blast07	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001701


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001701270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.11°C	-20.04°C	-20.04°C	-0.08°C
27-Apr-17 11:34	10.06°C	9.74°C	9.74°C	0.33°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
 Studio 19
 50-54 St Pauls Square
 Birmingham, West Midlands
 B3 1QS

CERTIFICATE NUMBER 00001706270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
EQY-Blast08	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001706


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001706270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:17	-20.12°C	-20.04°C	-20.04°C	-0.08°C
27-Apr-17 11:34	10.07°C	9.81°C	9.81°C	0.26°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001662270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
EQY-Blast09	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001662


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001662270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.11°C	-20.10°C	-20.10°C	-0.01°C
27-Apr-17 11:35	10.05°C	9.65°C	9.65°C	0.39°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001721270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
EQY-Blast10	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001721


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001721270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.12°C	-20.14°C	-20.14°C	0.03°C
27-Apr-17 11:35	10.05°C	9.75°C	9.75°C	0.30°C

The above calibration is applicable only to the relevant recorder and probe.

*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001669270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
EQY-Coldstore01	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001669


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001669270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:17	-20.12°C	-20.07°C	-20.07°C	-0.05°C
27-Apr-17 11:20	9.88°C	10.03°C	10.03°C	-0.16°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001672270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
LGN-Chill	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00002254
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1296
Probe Certificate Issue Date	12-Jul-16
Probe Certificate Reference Number	00002254110716
Equipment Code(s)	CI0001, CI0005

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001672


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001672270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:20	-19.99°C	-20.27°C	-20.27°C	0.28°C
27-Apr-17 11:00	9.93°C	9.92°C	9.92°C	0.00°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: **Jake Farrow**

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00003159270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
LSC-Blast01	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00002254
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1296
Probe Certificate Issue Date	12-Jul-16
Probe Certificate Reference Number	00002254110716
Equipment Code(s)	CI0001, CI0005

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00003159


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00003159270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 14:52	-20.17°C	-19.95°C	-19.95°C	-0.22°C
27-Apr-17 11:00	9.93°C	10.10°C	10.10°C	-0.17°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001704270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
LSC-Blast02	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00002254
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1296
Probe Certificate Issue Date	12-Jul-16
Probe Certificate Reference Number	00002254110716
Equipment Code(s)	CI0001, CI0005

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001704


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001704270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 14:43	-20.05°C	-20.03°C	-20.03°C	-0.02°C
27-Apr-17 11:01	9.92°C	9.99°C	9.99°C	-0.07°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001703270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
LSC-Blast03	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00002254
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1296
Probe Certificate Issue Date	12-Jul-16
Probe Certificate Reference Number	00002254110716
Equipment Code(s)	CI0001, CI0005

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001703


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001703270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 14:45	-20.09°C	-19.94°C	-19.94°C	-0.15°C
27-Apr-17 11:01	9.92°C	10.00°C	10.00°C	-0.08°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001702270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
LSC-Blast04	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00002254
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1296
Probe Certificate Issue Date	12-Jul-16
Probe Certificate Reference Number	00002254110716
Equipment Code(s)	CI0001, CI0005

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001702

Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001702270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 14:41	-20.03°C	-19.95°C	-19.95°C	-0.08°C
27-Apr-17 11:02	9.92°C	9.99°C	9.99°C	-0.07°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
 Studio 19
 50-54 St Pauls Square
 Birmingham, West Midlands
 B3 1QS

CERTIFICATE NUMBER 00001695270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
LSC-Blast05	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00002254
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1296
Probe Certificate Issue Date	12-Jul-16
Probe Certificate Reference Number	00002254110716
Equipment Code(s)	CI0001, CI0005

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001695


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001695270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-19.98°C	-20.36°C	-20.36°C	0.38°C
27-Apr-17 11:00	9.93°C	10.08°C	10.08°C	-0.15°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001694270417

Page 1 of 2

DATE OF ISSUE 27-Apr-17

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
LSC-Blast06	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00002254
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1296
Probe Certificate Issue Date	12-Jul-16
Probe Certificate Reference Number	00002254110716
Equipment Code(s)	CI0001, CI0005

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001694


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001694270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-19.98°C	-20.19°C	-20.19°C	0.21°C
27-Apr-17 11:00	9.93°C	10.10°C	10.10°C	-0.17°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001693270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
LSC-Blast07	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00002254
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1296
Probe Certificate Issue Date	12-Jul-16
Probe Certificate Reference Number	00002254110716
Equipment Code(s)	CI0001, CI0005

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001693


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001693270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-19.98°C	-20.40°C	-20.40°C	0.42°C
27-Apr-17 11:00	9.93°C	9.90°C	9.90°C	0.03°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001692270417

Page 1 of 2

DATE OF ISSUE 27-Apr-17

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
LSC-Blast08	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00002254
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1296
Probe Certificate Issue Date	12-Jul-16
Probe Certificate Reference Number	00002254110716
Equipment Code(s)	CI0001, CI0005

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001692


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001692270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-19.98°C	-20.19°C	-20.19°C	0.21°C
27-Apr-17 11:00	9.93°C	10.04°C	10.04°C	-0.11°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001691270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
LSC-Blast09	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00002254
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1296
Probe Certificate Issue Date	12-Jul-16
Probe Certificate Reference Number	00002254110716
Equipment Code(s)	CI0001, CI0005

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001691


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001691270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-19.98°C	-20.24°C	-20.24°C	0.25°C
27-Apr-17 10:07	10.02°C	10.00°C	10.00°C	0.02°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
 Studio 19
 50-54 St Pauls Square
 Birmingham, West Midlands
 B3 1QS

CERTIFICATE NUMBER 00001685270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
LSC-Blast10	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00002254
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1296
Probe Certificate Issue Date	12-Jul-16
Probe Certificate Reference Number	00002254110716
Equipment Code(s)	CI0001, CI0005

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001685


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001685270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-19.98°C	-20.30°C	-20.30°C	0.32°C
27-Apr-17 11:00	9.93°C	10.03°C	10.03°C	-0.10°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
 Studio 19
 50-54 St Pauls Square
 Birmingham, West Midlands
 B3 1QS

CERTIFICATE NUMBER 00001673270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
LSC-Coldstore01	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00002254
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1296
Probe Certificate Issue Date	12-Jul-16
Probe Certificate Reference Number	00002254110716
Equipment Code(s)	CI0001, CI0005

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001673


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001673270417

Page 2 of 2

Calibration Results


	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-19.98°C	-20.27°C	-20.27°C	0.28°C
27-Apr-17 11:01	9.93°C	10.00°C	10.00°C	-0.07°C

The above calibration is applicable only to the relevant recorder and probe.
*Error comparing the reference equipment with the equipment under test.
**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001674270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
LSC-Coldstore02	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00002254
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1296
Probe Certificate Issue Date	12-Jul-16
Probe Certificate Reference Number	00002254110716
Equipment Code(s)	CI0001, CI0005

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001674


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001674270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-19.98°C	-20.20°C	-20.20°C	0.22°C
27-Apr-17 11:00	9.93°C	9.94°C	9.94°C	-0.01°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001682270417

Page 1 of 2

DATE OF ISSUE 27-Apr-17

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
LSC-Factory	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00002254
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1296
Probe Certificate Issue Date	12-Jul-16
Probe Certificate Reference Number	00002254110716
Equipment Code(s)	CI0001, CI0005

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001682


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001682270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-19.98°C	-20.22°C	-20.22°C	0.24°C
27-Apr-17 11:00	9.93°C	10.10°C	10.10°C	-0.17°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
 Studio 19
 50-54 St Pauls Square
 Birmingham, West Midlands
 B3 1QS

CERTIFICATE NUMBER 00001666270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
RRD-Ambient	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001666


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001666270417

Page 2 of 2

Calibration Results


	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.11°C	-20.23°C	-20.23°C	0.12°C
27-Apr-17 11:32	10.06°C	9.64°C	9.64°C	0.42°C

The above calibration is applicable only to the relevant recorder and probe.
*Error comparing the reference equipment with the equipment under test.
**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001688270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
RRD-BigChill	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001688


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001688270417

Page 2 of 2

Calibration Results


	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:17	-20.12°C	-20.24°C	-20.24°C	0.12°C
27-Apr-17 11:20	9.88°C	9.85°C	9.85°C	0.03°C

The above calibration is applicable only to the relevant recorder and probe.
*Error comparing the reference equipment with the equipment under test.
**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001667270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
RRD-Coldstore01	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001667


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001667270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.11°C	-20.10°C	-20.10°C	-0.01°C
27-Apr-17 11:32	10.06°C	9.97°C	9.97°C	0.09°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
 Studio 19
 50-54 St Pauls Square
 Birmingham, West Midlands
 B3 1QS

CERTIFICATE NUMBER 00001700270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
RRD-Coldstore02	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001700


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001700270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:17	-20.12°C	-20.08°C	-20.08°C	-0.04°C
27-Apr-17 11:31	10.04°C	9.73°C	9.73°C	0.31°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001680270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
<input type="text" value="RRD-Coldstore03"/>	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001680


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001680270417

Page 2 of 2

Calibration Results


	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:17	-20.12°C	-20.11°C	-20.11°C	-0.01°C
27-Apr-17 11:32	10.06°C	9.87°C	9.87°C	0.19°C

The above calibration is applicable only to the relevant recorder and probe.
*Error comparing the reference equipment with the equipment under test.
**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001689270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
RRD-LongChill	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001689


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00001689270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.11°C	-20.25°C	-20.25°C	0.14°C
27-Apr-17 11:35	10.05°C	9.55°C	9.55°C	0.50°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001668270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
RRD-SmallChill	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001668

Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001668270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:17	-20.12°C	-20.23°C	-20.23°C	0.11°C
27-Apr-17 11:31	10.04°C	9.82°C	9.82°C	0.22°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00002093270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
USF-Cabinet01	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00002093


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00002093270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:17	-20.12°C	-20.15°C	-20.15°C	0.03°C
27-Apr-17 11:35	10.05°C	9.95°C	9.95°C	0.10°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
 Studio 19
 50-54 St Pauls Square
 Birmingham, West Midlands
 B3 1QS

CERTIFICATE NUMBER 00001699270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
USF-Cabinet02	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00001699


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00001699270417

Page 2 of 2

Calibration Results


	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:17	-20.12°C	-20.24°C	-20.24°C	0.12°C
27-Apr-17 11:32	10.04°C	9.94°C	9.94°C	0.10°C

The above calibration is applicable only to the relevant recorder and probe.
*Error comparing the reference equipment with the equipment under test.
**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00002092270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
USF-DispatchChill	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00002092


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00002092270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:17	-20.12°C	-20.21°C	-20.21°C	0.09°C
27-Apr-17 12:39	9.83°C	9.93°C	9.93°C	-0.10°C

The above calibration is applicable only to the relevant recorder and probe.


*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00002091270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
USF-Display	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00002091


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00002091270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:19	-20.11°C	-20.13°C	-20.13°C	0.02°C
27-Apr-17 12:39	9.84°C	9.87°C	9.87°C	-0.04°C

The above calibration is applicable only to the relevant recorder and probe.
*Error comparing the reference equipment with the equipment under test.
**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00002094270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
USF-HighRiskChill	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00002094


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER 00002094270417

Page 2 of 2

Calibration Results


	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:17	-20.12°C	-20.23°C	-20.23°C	0.11°C
27-Apr-17 11:32	10.05°C	9.76°C	9.76°C	0.29°C

The above calibration is applicable only to the relevant recorder and probe.
*Error comparing the reference equipment with the equipment under test.
**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
 Studio 19
 50-54 St Pauls Square
 Birmingham, West Midlands
 B3 1QS

CERTIFICATE NUMBER 00002095270417
DATE OF ISSUE 27-Apr-17

Page 1 of 2

Calibration Procedure	QP0000-1
-----------------------	----------

Customer	Lunar Freezing
Address	Lunar Cold Store Damhead Way Peterhead AB24 3GY
USF-IntakeChill	
Tel	
Fax	
Contact	Mark Cauvin
Order Number	5003325

Reference Equipment

Type	Notion Transmitter
Serial Number	00001377
Probe Type	PT 100 (Channel 1)
Probe Serial Number	PT1193
Probe Certificate Issue Date	28-Apr-16
Probe Certificate Reference Number	00001377280416
Equipment Code(s)	CI0001, CI0004

Unit Calibrated

Type	Notion Transmitter
Probe Type	Internal (MicroTemp)
Serial Number	00002095


Environmental Conditions

Temperature	16°C ±3°C	Humidity	50%RH ±20%RH
-------------	-----------	----------	--------------

Calibration Engineer	Darren
Approved Signatory	
Print Name	
Approved Signatory	
Sign Name	

Approver: Jake Farrow

Signature:



CERTIFICATE OF CALIBRATION

JTF Wireless Ltd
Studio 19
50-54 St Pauls Square
Birmingham, West Midlands
B3 1QS

CERTIFICATE NUMBER

00002095270417

Page 2 of 2

Calibration Results

	Reference	Unit Calibrated		
Date	Reference Mean	As Found**	UUT Mean	Error*
26-Apr-17 16:17	-20.12°C	-20.34°C	-20.34°C	0.22°C
27-Apr-17 12:39	9.84°C	9.78°C	9.78°C	0.05°C

The above calibration is applicable only to the relevant recorder and probe.

*Error comparing the reference equipment with the equipment under test.

**Device reading as found prior to any adjustment being made.

End of Document

Approver: Jake Farrow

Signature:

